

NGS High-Fidelity Pfu Buffer, 10x

For research or further manufacturing use only

Catalog No:	MDX038
Lot No:	CP051-B130320
Storage Conditions:	-20°C
Component Lot No:	PRB-424108B
Expiry date:	September 2026

Quality Control Parameters

Optimized for use with High-Fidelity Pfu (Cat# MDX003)

Analysis	Specification	Result
Functional	Fragment size of 1200bp is amplified from a specific dilution series of template DNA under defined conditions. The reactions are analysed using agarose gel electrophoresis (ethidium stained). A 1200bp band should be observed from dilution 1 to 9.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative:



X. Chen

Date: 20th August 2024

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01