

immunocard STAT!®



 Using the applicator stick of the diluent vial, transfer a small portion (5-6 mm diameter) of stool specimen into the sample diluent. Vortex 15 seconds.



Hold the vial vertically and gently tap on benchtop. Break the tip from vial using absorbent paper.



Dispense 4 drops into the sample port of the device. After 5 minutes, read result within 1 minute.

Interpretation of Results



POSITIVE



NEGATIVE



INVALID



USA/CORPORATE OFFICE 3471 River Hills Drive Cincinnati, Ohio 45244 USA Telephone: 513-271-3700 EMEA, MERIDIAN BIOSCIENCE EUROPE
Phone: +39 0331433636
E-mail: MBE-TechService@meridianbioscience.eu

Orders/Customer Service: 1-800-543-1980 Technical Support: 1-800-543-3858 Information Fax: 513-272-5432 Ordering Fax: 513-271-0124 meridianbioscience.com



© 2023 SN750220 REV. 02/23