

**Lyo-Ready™ Genotyping Direct
qPCR Blood**

Cat No: MDX128

Shipping conditions: -20 °C

Lot number: RD122L

Test Parameters

Analysis	Specification	Result
Functionality	<p>Quantitative PCR analysis amplifying a multiplex reaction of 2 genes from a dilution series of template under standard conditions.</p> <p>Relative activity is measured against a reference by a functional qPCR assay.</p> <p>Test Criteria:</p> <ul style="list-style-type: none"> - Test and reference amplification profiles must align; $\Delta Ct (\text{Test-Reference}) < 1$ 	Passed

Authorised by William Ferreira
Team Leader, R&D



Technical Support: mbi.tech@meridianlifescience.com

Bioline Reagents Ltd
UNITED KINGDOM

Meridian Life Science Inc
USA

Bioline GmbH
GERMANY

Bioline (Aus) Pty Ltd
AUSTRALIA

Bioline France
FRANCE

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

Tel: +1 901 382 8716
Fax: +1 901 382 0027

Tel: +49 (0)337 168 1229
Fax: +49 (0)337 168 1244

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10