

Certificate of Analysis

COA No: CA_BSM-0108

Version: 02

Air-Dryable™ Direct RNA/DNA qPCR Stool Mix, 4x

For research or further manufacturing use only

Catalog No:	MDX141
Lot No:	B128620
Storage Conditions:	-20°C
Component Lot No:	224306A
Expiry date:	July 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative real-time PCR analysis amplifying a target gene from a dilution series of mouse RNA under standard cycling conditions. Pass Criteria: Amplification profile of a 1:10 dilution must be consistent for the test and reference sample within ≤ 0.5 Cq difference. The end florescence of the 1:10 dilution must be consistent for the test and reference sample within ≤ 0.10 difference.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample. Pass Criteria: Amplification traces must overlay with the negative control.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 6.25 x 10 ⁻⁴ KU DNase I. Pass Criteria: No detectable degradation.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: 9.7 x 10 ⁻³ ng/µL RNase Pass Criteria: Test sample must show less RNase activity than the limit of detection.	Passed

QA / QC Representative:

7.24

J. Rahnenführer

Date: 27th June 2024

United Kingdom

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 <u>USA</u> Tel: +1 901.382.8716

<u>Germany</u>

Tel: +49 (0)3371 60222 00 Fax: +49 (0)3371 60222 01