

COA No: CA\_SUP-0003 (CA\_BSM-0108-2)

## Lyo-Ready<sup>™</sup> Direct RNA/DNA qPCR Stool Mix, 4x

Catalog No:	MDX143
Lot No:	B120810
Storage Conditions:	-20°C
Component Lot No:	223309A
Expiry date:	October 2025

For research or further manufacturing use only

## **Quality Control Parameters**

Analysis	Specification	Result
Functional	Quantitative real-time PCR analysis amplifying a target gene from a dilution series of mouse RNA under standard cycling conditions. <u>Pass Criteria:</u> Amplification profile of a 1:10 dilution must be consistent for the test and reference sample within ≤ 0.5 Cq difference.   The end florescence of the 1:10 dilution must be consistent for the test and reference sample within ≤ 0.10 difference.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample. <u>Pass Criteria</u> : Amplification traces must overlay with the negative control.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 6.25 x 10 <sup>-4</sup> KU DNase I. <u>Pass Criteria</u> : No detectable degradation.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: 9.7 x 10 <sup>-3</sup> ng/µL RNase <u>Pass Criteria</u> : Test sample must show less RNase activity than the limit of detection.	Passed

QA / QC Representative:

Andrew Galeeba-M

Date: 13<sup>th</sup> September 2023

United Kingdom

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 <u>USA</u> Tel: +1 901.382.8716 Fax: +1 901.382.0027 Germany

Tel: +49 (0)3371 60222 00 Fax: +49 (0)3371 60222 01

Info@meridianlifescience.com; www.meridianlifescience.com