

CERTIFICATE OF ANALYSIS

Important Note: Centrifuge before opening to ensure complete recovery of vial contents.

Catalog #: A32030H **Lot #:** 1A02822

Description: CEA Antigen (Calibrator Grade)
Carcinoembryonic Antigen (CEA) Calibrator Grade

Source: Cell Culture

Format: Calibrator, Liquid

Purification: > 50% pure (SDS-PAGE using glycoprotein stain).
Product is 0.2 µm filtered.

Concentration: 3.33mg/mL (CEA Elecsys)
Total Protein: 6-09 OD280nm Units

Buffer: Phosphate Buffered Saline, pH 7.4 ± 0.2

Preservative: 0.1% Sodium Azide


Applications: Specific methodologies have not been tested using this product.

Storage: Short-term store at 2–8°C. Long term store at -20°C. Avoid multiple freeze/thaw cycles.

Inactivation: Not Applicable

Safety Note(s): Refer to the appropriate Safety Data Sheets (SDS) for additional information.

Quality Signature:



31JAN2022

FOR RESEARCH OR FURTHER MANUFACTURING USE ONLY