

Inhibitor Tolerant RT-qPCR Mix, 4x


For research or further manufacturing use only

Catalog No:	MDX016
Lot No:	B124450
Storage Conditions:	-20°C
Component Lot No:	224201A
Expiry date:	February 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Amplification of a target gene from mouse Total RNA using a probe-based RT-qPCR assay under standard cycling conditions. <u>Pass Criteria:</u> Amplification profiles must be consistent for the test and reference sample within ± 1 Cq difference.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample. <u>Pass Criteria:</u> Amplification traces must overlay with the negative control.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 6.25×10^{-4} KU DNase I. <u>Pass Criteria:</u> No detectable degradation.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: 9.7×10^{-3} ng/ μ L RNase <u>Pass Criteria:</u> Test sample must show less RNase activity than the limit of detection.	Passed

QA / QC Representative:



Jan Rahnenführer

 Date: 17th January 2024

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01