

Certificate of Analysis

COA No: CA_BSM-0095-2

Version: 02

Lyo-Ready qPCR Mix, 2.6x

For research or further manufacturing use only

Catalog No:	MDX023	
Lot No:	EM030-B124480	
Storage Conditions:	-20°C	
Component Lot No:	LYN2-224101A	
Expiry date:	February 2026	

Quality Control Parameters

Glycerol-free formulation optimized for lyophilization and ambient temperature stabilization of qPCR assays (2.6x concentration)

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions.	
	Pass Criteria:	Passed
	C_{t} profiles must be consistent for the test and reference sample within \pm 0.5 Cq variance.	
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked.	
	Pass Criteria:	Passed
	Test sample must amplify in concordance with control sample.	
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis.	
	Limit of detection: 6.25 x 10 ⁻⁴ KU DNase I	Passed
	Pass Criteria:	
	No detectable degradation.	

QA / QC Representative:

Jan Rahnenführer

Date: 17th January 2024

United Kingdom

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 USA

Germany

Tel: +49 (0)3371 60222 00 Tel: +1 901.382.8716 Fax: +1 901.382.0027 Fax: +49 (0)3371 60222 01