

dUTP Mix 50mM

Suitable for Research and further Manufacturing Use


Catalog No:	MDX058
Lot No:	NU074-B124720
Storage Conditions:	-20°C
Component Lot No:	DUM-324101B
Expiry date:	February 2026

Quality Control Parameters

Ultra-pure (>99% by HPLC) dNTPs supplied as lithium salts

Analysis	Specification	Result
Functional	A 800bp human genomic DNA fragment is amplified with a dilution series of dUTPs, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed
Nicking Activity	Incubation of dUTP mix with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

QA / QC Representative:



J. Rahnenführer

 Date: 25th January 2024

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01