

## 4x Lyo-Ready™ Buffer

For research or further manufacturing use only

Catalog No:	MDX061
Lot No:	CP080-B105010
Storage Conditions:	-20°C
Component Lot No:	LB4-122202A
Expiry date:	March 2024

### Quality Control Parameters

Glycerol-free qPCR buffer formulated without excipients

Analysis	Specification	Result
Functional	<p>Quantitative RT-PCR analysis amplifying three targets in multiplex from a dilution series of mouse RNA under standard conditions. Ct profiles must be consistent for test and reference samples with a <math>\pm 0.5</math> Ct variance.</p> <p><u>Pass Criteria:</u></p> <p>The delta Rn of the amplification traces, for test and reference samples, must be within 10 %.</p>	Passed
DNA contamination	<p>Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked.</p> <p><u>Pass Criteria:</u></p> <p>Test sample must amplify in concordance with control sample.</p>	Passed
DNase contamination	<p>DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis.</p> <p>Limit of detection: <math>6.25 \times 10^{-4}</math> KU DNase I.</p> <p><u>Pass Criteria:</u></p> <p>No detectable degradation.</p>	Passed

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01

**Australia**

Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

# Certificate of Analysis

COA No: CA\_XBB-0063

Version: 06

RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase <u>Pass Criteria:</u> Test sample must show less RNase activity than the limit of detection.	Passed
---------------------	---	--------

QA / QC Representative:



Andrew Galeeba-M

Date: 11<sup>th</sup> February 2022

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01

**Australia**

Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763