

Fast qPCR Mix, 5x


For research or further manufacturing use only

Catalog No:	MDX072
Lot No:	EM100-B131660
Storage Conditions:	-20°C
Component Lot No:	FQP-224110A
Expiry date:	November 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Amplification of a target gene from mouse cDNA using a probe-based triplex qPCR assay under standard cycling conditions.</p> <p><u>Pass Criteria:</u></p> <p>Amplification profiles must be consistent for the test and reference sample within a 0.5 CT or 1.0 CT difference, depending on the amplicon.</p>	Passed
DNA contamination	<p>Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.</p> <p><u>Pass Criteria:</u></p> <p>Amplification traces must overlay with the negative control.</p>	Passed
DNase contamination	<p>DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis.</p> <p>Limit of detection: 6.25×10^{-4} KU DNase I.</p> <p><u>Pass Criteria:</u></p> <p>No detectable degradation.</p>	Passed

QA / QC Representative:



J. Rahnenführer

Date: 11th October 2024

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01