

BST Reaction Buffer, 10x

For research or further manufacturing use only

Catalog No:	MDX076
Lot No:	B122660
Storage Conditions:	-20°C
Component Lot No:	223111A
Expiry date:	December 2025

Quality Control Parameters

Analysis	Specification	Result
Functional	TTR (time-to-result) values obtained by Loop mediated amplification reaction of Test Sample vs Reference sample must be within 1.5 minutes	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.	Passed
DNase contamination	Incubation of a 1 Kb double stranded DNA fragment. Incubation for 4 hours at 37° C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must exhibit less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative:



Andrew Galeeba-M

 Date: 13th November 2023

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01