

## Air-Dryable™ Direct DNA qPCR Blood, 4x


For research or further manufacturing use only

Catalog No:	MDX092
Lot No:	B126280
Storage Conditions:	-20°C
Component Lot No:	424103A
Expiry date:	April 2026

### Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Quantitative real-time PCR analysis amplifying three target genes from a dilution series of mouse cDNA under standard cycling conditions.</p> <p><u>Pass Criteria:</u></p> <p>For amplification below Ct 30, Ct profiles must be consistent for test and reference samples with a <math>\pm 0.5</math> Ct variance.</p> <p>For amplification greater than Ct 30, Ct profiles must be consistent for test and reference samples with a <math>\pm 1</math> Ct variance.</p>	Passed
DNA contamination	<p>Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked.</p> <p><u>Pass Criteria:</u></p> <p>Test sample must amplify in concordance with control sample.</p>	Passed
DNase contamination	<p>DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis.</p> <p>Limit of detection: <math>6.25 \times 10^{-4}</math> KU DNase I.</p> <p><u>Pass Criteria:</u></p> <p>No detectable degradation.</p>	Passed
RNase contamination	<p>Quantitative PCR analysis with high and low RNase standards.</p> <p>Limit of detection: <math>9.7 \times 10^{-3}</math> ng/<math>\mu</math>L RNase</p> <p><u>Pass Criteria:</u></p> <p>Test sample must show less RNase activity than the limit of detection.</p>	Passed

QA / QC Representative:



J. Rahnenführer

 Date: 21<sup>st</sup> March 2024

United Kingdom

 Tel: +44 (0)20 8830 5300  
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716  
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00  
 Fax: +49 (0)3371 60222 01