

Certificate of Analysis

CA_BSM-0103-2

Version: 01

Air-Dryable™ Direct DNA qPCR Saliva, 4x

For research or further manufacturing use only

Catalog No:	MDX130	
Lot No:	B108910	
Storage Conditions:	-20°C	
Component Lot No:	122306A	
Expiry date:	July 2024	

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative real-time PCR analysis amplifying three target genes from a dilution series of mouse cDNA under standard cycling conditions. Pass Criteria: For amplification below Ct 30, Ct profiles must be consistent for test and reference samples with a ± 0.5 Ct variance. For amplification greater than Ct 30, Ct profiles must be consistent for test and reference samples with a ± 1 Ct variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Pass Criteria: Test sample must amplify in concordance with control sample.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 6.25 x 10 ⁻⁴ KU DNase I. Pass Criteria: No detectable degradation.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: $9.7 \times 10^{-3} \text{ ng/}\mu\text{L}$ RNase Pass Criteria: Test sample must show less RNase activity than the limit of detection.	Passed

QA / QC Representative:



Andrew Galeeba-M

Date: 24th June 2022

United Kingdom

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 USA

Tel: +1 901.382.8716 Fax: +1 901.382.0027 **Germany**

Tel: +49 (0)3371 60222 00 Fax: +49 (0)3371 60222 01 <u>Australia</u>

Tel: +61 (0)2 9209 4180 Fax: +61 (0)2 9209 4763