

Lyo-Ready™ Direct DNA LAMP Mix Saliva, 4x

Suitable for Research and further Manufacturing Use

Catalog No:	MDX134
Lot No:	B119780
Storage Conditions:	-20°C
Component Lot No:	323108A
Expiry date:	September 2025

Quality Control Parameters

Analysis	Specification	Result
Functional	TTR (time-to-result) values obtained by Loop mediated amplification reaction of Test Sample vs Reference sample must be within 1.5 minutes	Passed
DNA contamination	DNA contamination is measured by quantitative PCR on <i>E. coli</i> and mouse genomic DNA specific targets. <u>Pass Criteria:</u> Amplification traces must overlay with the negative control.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 2.5×10^{-3} U DNase I. <u>Pass Criteria:</u> No detectable degradation.	Passed

QA / QC Representative:



Andrew Galeeba-M

Date: 11th September 2023

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01