

**MDX135 Lyo-Ready™ Direct
RNA/DNA LAMP Saliva**

Cat. No. MDX135

Storage Conditions: -20 °C

Lot number: RD122C

Test Parameters

Analysis	Specification	Result
Functionality	<p>Performance is evaluated in a Loop-mediated Isothermal Amplification (LAMP) assay at 1000 input copies of template VLP, diluted 1:100 Zika virus and diluted 1:500 inactivated RSV virus in presence and absence of 20% human saliva</p> <p><u>Pass Criteria:</u></p> <p>Time-to-result (TTR) is consistent between test and reference mix.</p>	Passed

Authorised by Michele Amasio
Sr. Manager R&D



UNITED KINGDOM

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901 382 8716
Fax: +1 901 382 0027

GERMANY

Tel: +49 (0)337 168 1229
Fax: +49 (0)337 168 1244

AUSTRALIA

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

FRANCE

Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10