

Test Release Document

Released:
Michele Amasio

Version: 01

**Lyo-Ready™ Genotyping Direct
qPCR Stool, 4x**

Cat. No. MDX148

Storage Conditions: -20 °C

Lot number: RD123G

Test Parameters

Analysis	Specification	Result
Functionality	<p>Quantitative PCR analysis amplifying target from a dilution series of template under standard conditions.</p> <p>Relative activity is measured against a reference by a functional qPCR assay.</p> <p>Test Criteria: Test and reference amplification profiles must align; ΔCt (Test-Reference) < 1</p>	Passed

Authorised by William Ferreira
Team Leader R&D

