

COA No: CA_BSM-0111

Version: 01

Air-Dryable™ Direct DNA qPCR Urine, 4x

Catalog No:	MDX150	
Lot No:	B122420	
Storage Conditions:	-20°C	
Component Lot No:	323111A	
Expiry date:	December 2025	

For research or further manufacturing use only

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative real-time PCR analysis amplifying a target gene from dilutions of mouse cDNA under standard cycling conditions.	
	Pass Criteria:	- · ·
	Amplification profile of a 1:10 dilution must be consistent for the test and reference sample within \leq 1.0 Cq difference.	Passed
	The end florescence of the 1:10 dilution must be consistent for the test and reference sample within \leq 0.10 difference.	
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.	Passed
	Pass Criteria: Amplification traces must overlay with the negative control.	
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5 x 10 ⁻³ U DNase. Pass Criteria:	Passed
	No detectable degradation.	

QA / QC Representative:

Samantha Atkins

Som

Date: 13 November 2023

<u>United Kingdom</u> Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 <u>USA</u> Tel: +1 901.382.8716 Fax: +1 901.382.0027 <u>Germany</u>

Tel: +49 (0)3371 60222 00 Fax: +49 (0)3371 60222 01

Info@meridianlifescience.com; www.meridianlifescience.com