Lyo-Ready qPCR Mix, 2.6x Product Handling Guide		 Storage and stability: Lyo-Ready qPCR Mix, 2.6x is shipped on dry or blue ice. On arrival store at -20 °C for optimum stability. Repeated freeze/thaw cycles should be avoided. Thawing during transportation does not affect the product performance. Solutions should be mixed/equilibrated after each thawing to avoid phasing. Expiry: When stored under the recommended conditions and handled correctly, full activity of the kit is retained until the expiry date on the outer box label. 	
Batch No.: Concentration:	See vial 2.6x Store at –20 °C	Quality control: Bioline operates under ISO 13485 Quality Management System. Lyo-Ready qPCR Mix and its components are extensively tested for activity, processivity, efficiency, heat activation sensitivity, absence of nuclease contamination and absence of nucleic acid contamination.	
		Notes: This reagent has been manufactured under 13485 Quality Management System and is suitable for further manufacturing use as an IVD component.	

Description

Lyo-Ready qPCR Mix, 2.6x is a glycerol-free qPCR mix containing Taq polymerase, RNase inhibitor, reaction buffer, dNTP, MgCl₂ and lyoexcipients. In order to produce room temperature lyophilized qPCR reagents, assay specific primers and probes are added to Lyo-Ready qPCR Mix for subsequent lyophilization.

Kit components

Table 1

Component

Lyo-Ready qPCR Mix, 2.6x

Users Guidelines

Master mix preparation

Recommended reagent volumes per 20 μL qPCR mix are given in Table 2.

Table 2

Reagent	Volume
Lyo-Ready qPCR Mix, 2.6x	7.7 μL
Primer-Probe Mix, 20x	1 µL
Water	xμL
Total volume	Up to 20 µL

Dispense into reaction vials and immediately transfer into a freeze-dryer and run a suitable drying cycle.

For long-term storage at ambient temperatures, the lyophilized product from the freeze-dryer should be packaged under low relative humidity conditions.

Assay setup

Rehydrate the lyophilized qPCR mix with 20 μL template-containing solution and run qPCR.

The qPCR conditions in Table 3 are suitable for amplicons of up to 200 bp. These cycling parameters have been optimized for Lyo-Ready qPCR Mix, 2.6x on a number of platforms, however they can be varied to suit different machine-specific protocols.

Table 3

Step	Temperature	Time	Cycles
Polymerase activation	95 °C	2 min	1
Denaturation	95 °C	5 s	45
Annealing/Extension	60 °C	20 s	40

Technical Support

For any technical enquiries, please contact our Technical Support team via email at: mbi.tech@meridianlifescience.com

Bioline Reagents Ltd UNITED KINGDOM Bioline USA Inc. USA

Tel: +1 901 382 8716 Fax: +1 901 382 0027 Bioline GmbH GERMANY

Tel: +49 (0)337 160222 00 Fax: +49 (0)3371 60222 01 Bioline (Aust) Pty. Ltd AUSTRALIA

Tel: +61 (0)2 9209 4180 Fax: +61 (0)2 9209 4763 Bioline France FRANCE

Tel: +33 (0)1 42 56 04 40 Fax: +33 (0)9 70 06 62 10

Tel: +44 (0)20 8830 5300 Tel: +7 Fax: +44 (0)20 8452 2822 Fax: +

Website: www.meridianlifescience.com/molecular-dx-reagents email: info@meridianlifescience.com