Air-Dryable™ Direct RNA/DNA qPCR Blood **Product Handling Guide**

Shipping: Blue Ice MDX121 Catalog number:

Concentration: 4x

Store at -20 °C



See vial

Storage and stability

Air-Dryable™ Direct RNA/DNA gPCR Blood is shipped on blue ice. On arrival store at -20 °C for optimum stability. Repeated freeze/thaw cycles should be avoided. Thawing during transportation does not affect the product performance. Solutions should be mixed/equilibrated after each thawing to avoid phasing.

Expiry:

When stored under the recommended conditions and handled correctly, full activity of the kit is retained until the expiry date on the outer box label.

Safety precautions:

Read and understand the SDS (Safety Data Sheets) before handling the reagents. Hardcopies of the SDS will be provided with the first shipment, thereafter they will be available upon request.

Meridian operates under ISO 13485 Quality Management System. Air-Dryable™ Direct RNA/DNA qPCR Blood and its components are extensively tested for activity, processivity, efficiency, heat activation sensitivity, absence of nuclease contamination and absence of nucleic acid

Notes:

For research or further manufactured use only.

Description

Batch No .:

Air-Dryable™ Direct RNA/DNA qPCR Blood is a glycerol-free one tube formulation combining the lasts advances in buffer chemistry and PCR enhancers, together with an optimized concentration of antibody-mediated hot-start polymerase, reverse transcriptase, RNase Inhibitor, dNTPs and MgCl₂. and air-dry compatible excipients. Air-Dryable™ Direct RNA/DNA qPCR Blood has been developed for highly reproducible, accurate RNA and DNA target amplification, delivering excellent results in multiplex assays in presence of whole blood treated with different anticoagulants. In order to produce room temperature air-dried RT-qPCR reagents, assay specific primers and probes can be added to Air-Dryable™ Direct RNA/DNA gPCR Blood for subsequent airdrying.

Kit components

Table 1

Component

Air-Dryable™ Direct RNA/DNA qPCR Blood , 4x

Users Guidelines

The amount of inhibition tolerated by Air-Dryable™ Direct RNA/DNA qPCR Blood is variable depending on several factors, including assay design and sample quality. For this reason, an initial sample titration is recommended.

Master mix preparation

Recommended reagent volumes of Air-Dryable™ Direct RNA/DNA qPCR Blood and Primer-Probe Mix for air-drying are given in Table 2. Volumes are indicated per 20 µL final rehydrated reaction.

Table 2

Reagent	Volume
Air-Dryable™ Direct RNA/DNA qPCR Blood, 4x	5 μL
Primer–Probe Mix, 20x	1 μL*
Total Volume	6 μL

^{*} Primer and probe concentration needs to be optimised

Dispense into reaction vessels, immediately transfer into a convection oven and run a suitable drying cycle.

For long-term storage at ambient temperatures, the air-dried product from should be packaged with silica desiccant in a heat sealed pouch.

Assay setup

Rehydrate the air-dried qPCR master mix with 20 µL template-containing solution, vortex and run RT-qPCR.

The qPCR conditions in Table 3 are suitable for amplicons of up to 200 bp. These cycling parameters are compatible with Air-Dryable™ Direct RNA/DNA qPCR Blood on a number of platforms, however they can be varied to suit different detection assay and machine-specific protocols.

Table 3

Step	Temperature	Time	Cycles
Reverse transcription*	50 °C	10 min	1
Polymerase activation	95 °C	2 min	1
Denaturation	95 °C	5 s	45
Annealing/Extension*	60 °C	20 s	45

^{*}When multiplexing, the reverse transcription reaction time can be extended up to 20 minutes and the annealing/extension time can be extended up to 60 seconds and/or the annealing/extension temperature can be increased up to 65° C.

Associated Products

Air-Dryable™ Direct qPCR Blood	MDX092
Air-Dryable™ 1-Step RT-qPCR Mix	MDX095
Air-Dryable™ Direct RNA/DNA qPCR Saliva	MDX131
Air-Dryable™ Direct RNA/DNA qPCR Stool	MDX141

Technical Support

For any technical enquiries, please contact our Technical Support team via email at: mbi.tech@meridianlifescience.com

Meridian Life Science Inc.

Tel: +1 901 382 8716

Fax: +1 901 382 0027